

Manual Supplement

Manual Title:	1750 Calibration	Supplement Issue:	2
Print Date:	November 2007	Issue Date:	6/13
Revision/Date:	1, 12/08	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 10, under *Rogowski Current Verification*,

Change: 6. Press the **WAVE MENUS** softkey and ensure the phase angle is -90 °.

To: 6. Press the **WAVE MENUS** softkey and ensure the auxφNRM phase angle is -90 °.


On page 11, under *CT Current Verification*,

Change: 5. Press the **WAVE MENUS** softkey and ensure the phase angle is 0.0 °.

To: 5. Press the **WAVE MENUS** softkey and ensure the auxφNRM phase angle is 0.0 °.

Change #2, 65922

On page 3, replace the CAT III and CAT IV with the following and add:

	Conforms to relevant South Korean EMC Standards
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.

On page 3, under **General Specifications** remove the **Conformance** section.

On page 6, under **Environmental and Safety Specifications** remove **Standards Conformance, EMC and Safety** and replace with:

- Safety..... IEC 61010-1: Installation 300 V CAT II, Measurement 600 V CAT IV / 1000 V CAT III, Pollution Degree 2
- Altitude 2000 m
- IP Rating IEC 60529: IP30
- Electromagnetic Environment IEC 61326-1: Basic
- Electromagnetic Compatibility..... Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.