

Manual Supplement

Manual Title: 1750 Getting Started Supplement Issue: **2**
Print Date: October 2006 Issue Date: 6/13
Part Number: 2386767 Page Count: 1
Revision/Date: 2, 7/11

This supplement contains information necessary
to ensure the accuracy of the above manual.

Change #1

On page 5, Table 2 add:

CS1750/1760 Case Polyester, for Fluke 1750 & 1760	2673962
---	---------

On page 5, Table 2:

Change:

Personal Digital Assistant (PDA) with USB cable and accessories	2386780
---	---------

To:

Archos A43 Internet Tablet (pre-installed with Fluke Power View), stylus and pouch	Contact Fluke
--	---------------

On page 5, Table 2, remove:


Sets of international ac power plug adapters for the PDA charger	2583479
--	---------

On page 9, delete the ***Installing the Software*** and ***Installing Power View on the PDA*** sections and replace with:

For complete instructions on how to install the **Archos 43 Internet Tablet**, go to the **Power Quality Software Downloads** page at www.Fluke.com.

Change #2

On page 3, replace the CAT III and CAT IV with the following and add:

	Conforms to relevant South Korean EMC Standards
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.